

बिड दस्तावेज़ / Bid Document

| बिड विवरण/Bid Details | |
|--|--|
| बिड बंद होने की तारीख/समय /Bid End Date/Time | 12-11-2025 16:00:00 |
| बिड खुलने की तारीख/समय /Bid Opening Date/Time | 12-11-2025 16:30:00 |
| बिड पेशकश वैधता (बंद होने की तारीख से)/Bid Offer Validity (From End Date) | 60 (Days) |
| मंत्रालय/राज्य का नाम/Ministry/State Name | Ministry Of Health And Family Welfare |
| विभाग का नाम/Department Name | Department Of Health And Family Welfare |
| संगठन का नाम/Organisation Name | Regional Institute Of Paramedical & Nursing Sciences(ripans) |
| कार्यालय का नाम/Office Name | Aizawl, Mizoram |
| कुल मात्रा/Total Quantity | 1 |
| वस्तु श्रेणी /Item Category | 5 Part Automated Hematology Analyser (V2) (Q2) |
| बिडर का न्यूनतम औसत वार्षिक टर्नओवर (3 वर्षों का) /Minimum Average Annual Turnover of the bidder (For 3 Years) | 3 Lakh (s) |
| मूल उपकरण निर्माता का औसत टर्नओवर (गत 3 वर्षों का)/OEM Average Turnover (Last 3 Years) | 6 Lakh (s) |
| उन्हीं/समान सेवा के लिए अपेक्षित विगत अनुभव के वर्ष/Years of Past Experience Required for same/similar service | 1 Year (s) |
| वर्षों के अनुभव एवं टर्नओवर से एमएसई को छूट प्राप्त है / MSE Exemption for Years Of Experience and Turnover | Yes Complete |
| स्टार्टअप के लिए अनुभव के वर्षों और टर्नओवर से छूट प्रदान की गई है /Startup Exemption for Years of Experience and Turnover | No |
| विक्रेता से मांगे गए दस्तावेज़/Document required from seller | Experience Criteria,Past Performance,Bidder Turnover,OEM Authorization Certificate,OEM Annual Turnover *In case any bidder is seeking exemption from Experience / Turnover Criteria, the supporting documents to prove his eligibility for exemption must be uploaded for evaluation by the buyer |

| बिड विवरण/Bid Details | |
|--|--|
| क्या आप निविदाकारों द्वारा अपलोड किए गए दस्तावेजों को निविदा में भाग लेने वाले सभी निविदाकारों को दिखाना चाहते हैं? संदर्भ मेनू है/Do you want to show documents uploaded by bidders to all bidders participated in bid? | Yes (Documents submitted as part of a clarification or representation during the tender/bid process will also be displayed to other participated bidders after log in) |
| बिड लगाने की समय सीमा स्वतः नहीं बढ़ाने के लिए आवश्यक बिड की संख्या। / Minimum number of bids required to disable automatic bid extension | 1 |
| दिनों की संख्या, जिनके लिए बिड लगाने की समय-सीमा बढ़ाई जाएगी। / Number of days for which Bid would be auto-extended | 3 |
| ऑटो एक्स्टेंशन अधिकतम कितनी बार किया जाना है। / Number of Auto Extension count | 1 |
| विगत प्रदर्शन /Past Performance | 10 % |
| बिड से रिवर्स नीलामी सक्रिय किया जाना है। / Bid to RA enabled | No |
| लागू आरसीएम/RCM Applicable | Yes |
| व्यापक रखरखाव शुल्क आवश्यक / Comprehensive Maintenance Charges Required | Yes |
| बिड का प्रकार/Type of Bid | Two Packet Bid |
| तकनीकी मूल्यांकन के दौरान तकनीकी स्पष्टीकरण हेतु अनुमत समय /Time allowed for Technical Clarifications during technical evaluation | 2 Days |
| निरीक्षण आवश्यक (सूचीबद्ध निरीक्षण प्राधिकरण /जेम के साथ पूर्व पंजीकृत एजेंसियों द्वारा)/Inspection Required (By Empanelled Inspection Authority / Agencies pre-registered with GeM) | No |
| मूल्यांकन पद्धति/Evaluation Method | Total value wise evaluation |
| मध्यस्थता खंड/Arbitration Clause | No |
| सुलह खंड/Mediation Clause | No |

ईएमडी विवरण/EMD Detail

| | |
|-----------------------------|----------------|
| एडवाइजरी बैंक/Advisory Bank | Bank Of Baroda |
| ईएमडी राशि/EMD Amount | 19440 |

ईपीबीजी विवरण /ePBG Detail

| | |
|-----------------------------|----------------|
| एडवाइजरी बैंक/Advisory Bank | Bank Of Baroda |
|-----------------------------|----------------|

| | |
|--|------|
| ईपीबीजी प्रतिशत (%) / ePBG Percentage (%) | 3.00 |
| ईपीबीजी की आवश्यक अवधि (माह) / Duration of ePBG required (Months). | 13 |

(a). जेम की शर्तों के अनुसार ईएमडी छूट के इच्छुक बिडर को संबंधित केटेगरी के लिए बिड के साथ वैध समर्थित दस्तावेज प्रस्तुत करने हैं। एमएसई केटेगरी के अंतर्गत केवल वस्तुओं के लिए विनिर्माता तथा सेवाओं के लिए सेवा प्रदाता ईएमडी से छूट के पात्र हैं। व्यापारियों को इस नीति के दायरे से बाहर रखा गया है। / EMD EXEMPTION: The bidder seeking EMD exemption, must submit the valid supporting document for the relevant category as per GeM GTC with the bid. Under MSE category, only manufacturers for goods and Service Providers for Services are eligible for exemption from EMD. Traders are excluded from the purview of this Policy.

(b). ईएमडी और संपादन जमानत राशि, जहां यह लागू होती है, लाभार्थी के पक्ष में होनी चाहिए। / EMD & Performance security should be in favour of Beneficiary, wherever it is applicable.

लाभार्थी /Beneficiary :

Director

Aizawl, Mizoram, Department of Health and Family Welfare, Regional Institute of Paramedical & Nursing Sciences(RIPANS), Ministry of Health and Family Welfare
(Dr.sanjay D. Sawant)

बोली विभाजन लागू नहीं किया गया/ Bid splitting not applied.

एमआईआई खरीद वरीयता / MII Purchase Preference

| | |
|---|---|
| एमआईआई खरीद वरीयता / MII Purchase Preference | Yes |
| मेक इन इंडिया विक्रेताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में है। / Purchase Preference to MII sellers available upto price within $L1+X\%$ | 20 |
| मेक इन इंडिया खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MII purchase preference | 50 |
| सार्वजनिक खरीद (मेक-इन-इंडिया को प्राथमिकता) आदेश 2017 के अनुसार केवल क्लास 1/क्लास 2 के स्थानीय आपूर्तिकर्ताओं को ही भागीदारी की अनुमति है दिनांक 16.09.2020 (समय-समय पर संशोधित एवं लागू) / Allow participation only from Class 1/Class 2 local suppliers as per the Public procurement(Preference to Make-in-india) order 2017 date 16.09.2020(as amended and applicable time to time) | Yes, in compliance with the MII ORDER : DPIIT Order(as amended and applicable time to time) |

एमएसई खरीद वरीयता/MSE Purchase Preference

| | |
|---|-----|
| एमएसई खरीद वरीयता/MSE Purchase Preference | Yes |
| सूक्ष्म और लघु उद्यम मूल उपकरण निर्माताओं को खरीद में प्राथमिकता, यदि उनका मूल्य $L1+X\%$ तक की सीमा में हो। / Purchase Preference to MSE OEMs available upto price within $L1+X\%$ | 15 |

| | |
|---|----|
| सूक्ष्म और लघु उद्यम को खरीद में प्राथमिकता के लिए बिड की मात्रा का अधिकतम प्रतिशत / Maximum Percentage of Bid quantity for MSE purchase preference | 25 |
|---|----|

1. If the bidder is a Micro or Small Enterprise as per latest orders issued by Ministry of MSME, the bidder shall be exempted from the eligibility criteria of "Experience Criteria" as defined above subject to meeting of quality and technical specifications. The bidder seeking exemption from Experience Criteria, shall upload the supporting documents to prove his eligibility for exemption.
2. If the bidder is a Micro or Small Enterprise (MSE) as per latest orders issued by Ministry of MSME, the bidder shall be exempted from the eligibility criteria of "Bidder Turnover" as defined above subject to meeting of quality and technical specifications. If the bidder itself is MSE OEM of the offered products, it would be exempted from the "OEM Average Turnover" criteria also subject to meeting of quality and technical specifications. The bidder seeking exemption from Turnover, shall upload the supporting documents to prove his eligibility for exemption.
3. The minimum average annual financial turnover of the bidder during the last three years, ending on 31st March of the previous financial year, should be as indicated above in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the bidder is less than 3-year-old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
4. Experience Criteria: In respect of the filter applied for experience criteria, the Bidder or its OEM of the product offered in the bid {themselves or through reseller(s)} should have regularly, manufactured and supplied same or similar Category Products to any Central / State Govt Organization / PSU for number of Financial years as indicated above in the bid document before the bid opening date. Copies of relevant contracts and delivery acceptance certificates like CRAC to be submitted along with bid in support of having supplied some quantity during each of the Financial year. In case of bunch bids, the category of primary product having highest value should meet this criterion.
5. OEM Turn Over Criteria: The minimum average annual financial turnover of the OEM of the offered product during the last three years, ending on 31st March of the previous financial year, should be as indicated in the bid document. Documentary evidence in the form of certified Audited Balance Sheets of relevant periods or a certificate from the Chartered Accountant / Cost Accountant indicating the turnover details for the relevant period shall be uploaded with the bid. In case the date of constitution / incorporation of the OEM is less than 3 year old, the average turnover in respect of the completed financial years after the date of constitution shall be taken into account for this criteria.
6. Preference to Make In India products (For bids < 200 Crore):Preference shall be given to Class 1 local supplier as defined in public procurement (Preference to Make in India), Order 2017 as amended from time to time and its subsequent Orders/Notifications issued by concerned Nodal Ministry for specific Goods/Products. The minimum local content to qualify as a Class 1 local supplier is denoted in the bid document. If the bidder wants to avail the Purchase preference, the bidder must upload a certificate from the OEM regarding the percentage of the local content and the details of locations at which the local value addition is made along with their bid, failing which no purchase preference shall be granted. In case the bid value is more than Rs 10 Crore, the declaration relating to percentage of local content shall be certified by the statutory auditor or cost auditor, if the OEM is a company and by a practicing cost accountant or a chartered accountant for OEMs other than companies as per the Public Procurement (preference to Make-in -India) order 2017 dated 04.06.2020. Only Class-I and Class-II Local suppliers as per MII order dated 4.6.2020 will be eligible to bid. Non - Local suppliers as per MII order dated 04.06.2020 are not eligible to participate. However, eligible micro and small enterprises will be allowed to participate .The buyers are advised to refer the OM No.F.1/4/2021-PPD dated 18.05.2023.
- OM No.1_4_2021_PPD_dated_18.05.2023 for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017.
7. Purchase preference will be given to MSEs having valid Udyam Registration and whose credentials are validated online through Udyam Registration portal as defined in Public Procurement Policy for Micro and Small Enterprises (MSEs) Order, 2012 dated 23.03.2012 issued by Ministry of Micro, Small and Medium Enterprises and its subsequent Orders/Notifications issued by concerned Ministry. If the bidder wants to avail themselves of the Purchase preference, the bidder must be the manufacturer / OEM of the offered product on GeM. Traders are excluded from the purview of Public Procurement Policy for Micro and Small Enterprises and hence resellers offering products manufactured by some other OEM are not eligible for any purchase preference. In respect of bid for Services, the bidder must be the Service provider of the offered Service. Relevant documentary evidence in this regard shall be uploaded along with the bid in respect of the offered product or service and Buyer will decide eligibility for purchase preference based on documentary evidence submitted, while evaluating the bid. If L-1 is

not an MSE and MSE Seller (s) has / have quoted price within L-1+ 15% (Selected by Buyer) of margin of purchase preference /price band defined in relevant policy, such MSE Seller shall be given opportunity to match L-1 price and contract will be awarded for 25% (selected by Buyer) percentage of total quantity. The buyers are advised to refer the OM No. F.1/4/2021-PPD dated 18.05.2023 [OM No.1 4 2021 PPD dated 18.05.2023](#) for compliance of Concurrent application of Public Procurement Policy for Micro and Small Enterprises Order, 2012 and Public Procurement (Preference to Make in India) Order, 2017. Benefits of MSE will be allowed only if seller is validated on-line in GeM profile as well as validated and approved by Buyer after evaluation of documents submitted.

8. Past Performance: The Bidder or its OEM {themselves or through re-seller(s)} should have supplied same or similar Category Products for 10% of bid quantity, in at least one of the last three Financial years before the bid opening date to any Central / State Govt Organization / PSU. Copies of relevant contracts (proving supply of cumulative order quantity in any one financial year) to be submitted along with bid in support of quantity supplied in the relevant Financial year. In case of bunch bids, the category related to primary product having highest bid value should meet this criterion.

जीएसटी की धारा 9(3)।/Section 9(3) Of GST

Where ever RCM is applicable, sellers (Regular GST registered seller who opted out of FCM , unregistered seller, seller registered under composition scheme)will be forced to put Zero GST and GST cess in their bids. Buyer will have liability of paying the GST and GST cess to the government on the specified rate mentioned by them in this Bid.

जीएसटी की धारा 9(3)। / Section 9(3) Of GST

Where ever RCM is applicable, sellers (Regular GST registered seller who opted out of FCM , unregistered seller, seller registered under composition scheme)will be forced to put Zero GST and GST cess in their bids. Buyer will have liability of paying the GST and GST cess to the government on the specified rate mentioned by them in this Bid.

If the buyer has mentioned MSE purchase preference in ATC then service provider is required to upload necessary documents for MSE purchase preference for verification by the buyer during evaluation.

5 Part Automated Hematology Analyser (V2) (1 pieces)

(क्रमशः श्रेणी 1 और श्रेणी 2 के स्थानीय आपूर्तिकर्ता के रूप में अर्हता प्राप्त करने के लिए आवश्यक/Minimum 50% and 20% Local Content required for qualifying as Class 1 and Class 2 Local Supplier respectively)

तकनीकी विशिष्टियाँ /Technical Specifications

* जेम केटेगरी विशिष्टि के अनुसार / As per GeM Category Specification

| विवरण/Specification | विशिष्टि का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|---|---|
| GENERAL | Product Description | 5 Part Automated Hematology Analyser |
| PRODUCT INFORMATION | Type of Configuration | Benchtop |
| | Type of system offered | Closed system |
| | Type of automation | Fully Automatic |
| | Equipment should have automatic start up, shut down and sample analysis | Yes |
| | Analysis principle | Based on principle of counting and sizing, Laser based scatter analysis |
| | Type of cell counting | 5 part WBC differential without Retic and NRBC enumeration capability |

| विवरण/Specification | विवरण/Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|---|---|
| | Should have multi channel analysis for better resolution and reproductability | Yes |
| | Analysis available | WBC, Lympho#, Lympho%, Neutrophil#, Neutrophil %, Basophil#, Basophil%, Eosinophil #, Eosinophil%, Monocyte#, Monocyte%, RBC, HGB, MCV, PCV, MCH, MCHC, RDW-SD, RDW CV, PLT, MPV, PDW, PLT, PLCR, PLCC, HCT, PCT, RET#, RET%, NRBC# |
| | NRBC% | No |
| | Availability of body fluid (CSF, synovial, serous) parameters like WBC-BF, RBC-BF, MN#, MN%, PMN%, PMN#, TC-BF# | No |
| | IPF analysis | No |
| | PLT-0 Analysis | No, Yes |
| | PLT-F Analysis | No, Yes |
| | IG % analysis | Yes |
| | IG# analysis | No, Yes |
| | RET-HE analysis | No |
| | HFR analysis | No |
| | LFR analysis | No |
| | MFR analysis | No |
| | IRFanalysis | No |
| | Discrete analysis modes available | CBC,CBC+DIFF |
| | Analysis method for WBC | Tri-angle Laser flow cytometry |
| | Method for platelet measurement | Electrical impedance |
| | RBC Measurement method | Electrical Impedance |
| | Hb measurement | Photometric/direct cellular measurement, Cyanide free colorimetry |
| | Retic measurement | Not provided |
| | Number of sample racks to cater to different tube sizes | 1 |
| | Maximum sample Aspiration volume needed in all modes | 20µl or less |
| | Minimum sample volume required in all modes | 20µl or less |
| | Throughput capacity in CBC/Differential | 50-60 |

| विवरण/Specification | विवरण का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|-----------------------------|---|---|
| | Throughput capacity in Retic mode | NA |
| | Type of modes of sample running | Open vial |
| | Linearity of Platelet | 0 to 3000 * 10 ³ cells/ micro litres |
| | RBC Linearity | 0 to 8 x 10 ⁶ per micro litre or more |
| | Hemoglobin linearity | 0 to 25 gm per litre |
| | Retic linearity | NA |
| | WBC linearity | 0 to 300 * 1000 cells/ micro litres |
| | Directly measures MCV | Yes |
| | Time taken by the analyser to produce the test results(Analysis time) in seconds | 40-60 |
| | Availability of Auto dilution | No, Yes |
| | Clot detection facility | No, Yes |
| | Automatic probe wipe | Yes |
| | Extended Analyses for cytopenic samples | Yes |
| | Quality assurance system with calibration and controls | Yes |
| | Number of quality control programs | Atleast 3 |
| | Type of Calibration | Both manual and automatic |
| | Direct aspiration for capillary blood from finger prick | No |
| | Floating discriminator for platelets and RBC counting for reliable RBC and PLT data | Yes |
| | Separate diluting nozzles for RBC and WBC | Yes |
| | Double bathing mechanism | Yes |
| | Upgradable to integrate with attachment | No |
| Data Management and Display | Type of data management | In-built system |
| | Display | TFT, LCD, LED Or higher |
| | PC hard disk | NA |
| | Inbuilt monitor size in inches | >5 |
| | Processor of PC Provided with system | NA (If PC not provided) |

| विवरण/Specification | विवरणि का नाम /Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|-----------------------|--|---|
| HARDWARE REQUIREMENTS | RAM | NA (If PC not provided) |
| | HDD | NA (If PC not provided) |
| | PC Monitor size if PC provided externally | NA (for Inbuilt based) |
| | The processor and RAM of the board system should be latest version | Yes |
| | HIS/LIS Interface | HL7 |
| | Type of external storage | USB |
| | Number of USB Port provided | 4 |
| | Type of printer unit | External |
| | Printer type | Black and White Laser Printer |
| | Display and print provided | Scatter plot and histograms |
| | Auto loader facility | Yes |
| | Facility for user defined flagging | Yes |
| | Database capability of storing sets of results and graphics | ? 50000 |
| | LJ Plot facility | Yes |
| | Delta check for cumulative review | No |
| | On-line QC option | Yes |
| | Patient moving average | Yes |
| | QC File management | Yes |
| | Facility for workload recording | Yes |
| | Flagging in event of unacceptable control data | Yes |
| | Ability to transmit results to host computer | Yes |
| | Type of user Interface or data entry | Touchscreen, Handheld barcode reader facility, Manual |
| | Have auto cleaning function in the analyser's software | Yes |
| POWER REQUIREMENTS | Type of power supply | 100-240 VAC,50-60 Hz |
| | Power Backup facility | UPS |
| | Type of UPS | Offline |
| | Rating of UPS in KVA | 0.5 |
| | Back up time in minutes | 30, 60, 90, 120 Or higher (minute) |

| विवरण/Specification | विवरण/Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|--|--|--|
| | Availability of Sufficient Number of Quality control files which store 100 or more XB Analysis | Yes |
| ACCESSORIES, SPARE PARTS AND CONSUMABLES | Offered equipment unit to be supplied with sufficient consumables (with at least 2/3rd of total shelf life) required for, sufficient to carry out haematological testing of samples" | 500 |
| | Should have on board reagent facility and automatic reagent inventory management | Yes |
| | Net work integration with lab information system feature | Yes |
| | Operating temperature and humidity | Capable of operating continuously in ambient temperature of 10 to 35 deg C and relative humidity of 15 to 85% in ideal circumstances |
| CERTIFICATION AND REPORTS | Compliance to Medical Device Rules (MDR) 2017 as amended till date | Yes |
| | Availability of valid drug license for the product issued from the competent authority defined under Drugs and Cosmetic Act 1940 and Rules made there under as amended till date | Yes |
| | Manufacturing unit certification | ISO:13485 (Latest) |
| | Availability of Test Report for product as per Medical Device Rules (MDR) 2017 as amended till date | Yes |
| | Electrical Safety Standards | IEC/EN 60601-1 or equivalent BIS Standard |
| | Submission of all necessary certifications, licenses and test reports to the buyer at the time of bid submission and/or along with supplies as per buyer requirement | Yes |
| WARRANTY | Warranty in years (Option of comprehensive warranty is available through bidding only, which if opted will supersede normal warranty in the catalogue) | 3, 5 Or higher (year) |
| MISCELLANEOUS REQUIREMENTS | OEM/Reseller shall ensure uninterrupted availability of all spares for 10 years | Yes |

| विवरण/Specification | विवरण/Specification Name | बिड के लिए आवश्यक अनुमत मूल्य /Bid Requirement (Allowed Values) |
|---------------------|---|---|
| | Availability of toll free facility for technical support maintained by OEM or authorized agencies | Yes |
| | User/Technical/Maintenance manuals to be supplied in English in hard and soft copy | Yes |
| | Details of equipments and procedures required for local calibration and routine maintenance to be supplied and advanced maintenance task documentation also to be furnished | Yes |
| | List of important spares and accessories, with their part numbers to be supplied to the buyer at the time of supplying the equipment | Yes |
| | Installation and Demonstration of equipment and training to be provided after completing supplies before acceptance | Yes |
| | The Principal Manufacturer must have direct Presence/approved service center In India | Yes |
| | Calibration certificates as per NABH requirement | Yes |
| | Time to attend breakdown calls | within 48 hrs |

| व्यापक रखरखाव / Comprehensive Maintenance | |
|--|--------|
| Warranty of required product | 3 Year |
| Comprehensive Maintenance Duration (Post Warranty) | 5 Year |
| GST as per RCM | 18% |
| GST Cess 1 as per RCM | 1% |
| Optional RCM | No |

*Warranty displayed under the AMC/CMC Details section will supersede the warranty displayed under the catalog specification

इनपुट कर क्रेडिट(आईटीसी) तथा रिवर्स प्रभार (आरसीएम)/Input Tax Credit(ITC) and Reverse Charge(RCM) Details

| जीएसटी पर इनपुट कर क्रेडिट /ITC on GST | जीएसटी उपकर कर क्रेडिट /ITC on GST Cess | लागू आरसीएम/RCM Applicable | रिवर्स प्रभार के अनुसार जीएसटी/GST as per RCM | रिवर्स प्रभार के अनुसार जीएसटी उपकार 1 /GST Cess 1 as per RCM | GST Cess 2 as per RCM | वैकल्पिक रिवर्स प्रभार /Optional RCM |
|--|---|----------------------------|---|---|-----------------------|--------------------------------------|
| NA | NA | Yes | 18% | 1% | 2 (INR) - Per Unit | Yes |

परेषिती/रिपोर्टिंग अधिकारी तथा मात्रा/Consignees/Reporting Officer and Quantity

| क्र.सं./S.N o. | परेषिती/रिपोर्टिंग अधिकारी /Consignee Reporting/Officer | पता/Address | मात्रा /Quantity | डिलीवरी के दिन/Delivery Days |
|----------------|---|---------------|------------------|------------------------------|
| 1 | Albert Lalrinngheta | 796017,RIPANS | 1 | 45 |

Special terms and conditions-Version:1 effective from 09-08-2023 for category 5 Part Automated Hematology Analyser (V2)

1. All Provisions of Drugs and Cosmetics Act, 1940 and Rules (including Medical Device Rule 2017) made there under as amended till date will always be applicable. This will include all notifications issued by Central Drugs Standard Control Organization (CDSCO), Ministry of Health & Family Welfare (MoHFW) and Department of Pharmaceuticals (DOP), Ministry of Chemicals & Fertilizers time to time in this regard.
2. The sellers are registered on GeM based on self-declaration of valid Drug License, product certification, test reports etc. However, buyers must check and validate the details at their end for all applicable licenses and certifications e.g., validity and authenticity/genuineness of drug license, product certification, manufacturer certification/licenses, test reports etc.
3. In case of authorized resellers/distributors, it will be the legal & regulatory liability of the manufacturer to ensure that their resellers/distributors are operating in compliance with all relevant laws and regulations and are properly licensed to sell the manufacturer's products, including verifying the validity and authenticity of drug license held by them.
4. The price offered by the seller/bidder shall not, in any case exceed the DPCO/NPPA controlled price or price fixed by State Government, if any. The seller must reduce the prices if there is any reduction in DPCO/NPPA ceiling price or price fixed by State Government, if any.
5. Any other Terms and Conditions which is not included or at variance with the conditions specified in STC/GTC, may be added by the buyer through Additional Terms and Conditions (ATC) in the bid to ensure items are procured from authentic/validated source with appropriate and applicable quality. The above terms and conditions are in reverse order of precedence i.e. ATC shall supersede specific STC which shall supersede General Terms and Conditions ("GTC"), whenever there are any conflicting provisions.
6. **Comprehensive warranty:** Comprehensive warranty shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares. During the warranty period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months. warranty shall not be including the consumables. Further there will be 98% uptime warranty during warranty period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend warranty period by double the downtime period.
7. **Service centres:** Details of Service outlets in India to render services for equipment to be furnished to buyer/consignees with complete address, telephone numbers, e mails etc at time of making the supplies. It shall be the responsibility of seller to ensure that authorized service centres are available to cater to the areas where supplies are made within reasonable distance from where the service calls can be handled. Details of toll-free numbers for service call and online registration of service requests also to be provided buyer/consignee at the time of supplies.

8. **Source of supply:** It shall be responsibility of seller to provide Documents regarding source of equipments such as copy of Performa invoice or any other documents to establish that the products supplied are manufactured by OEM indicated and sourced from them.
9. **Packing and Marking:** Medical equipments being very delicate and sensitive packing for the goods should be strong and durable enough to withstand transit including transhipment (if any), rough handling, open storage etc. without any damage, deterioration etc. .The size, weights and volumes of the packing cases, remoteness of the final destination of the goods, availability or otherwise of transport and handling facilities at all points during transit up to final destination,. Quality of packing, the manner of marking within & outside the packages and provision of accompanying documentation shall take into consideration the type of medical equipments being supplied. The accessories shall be suitably labelled and packed. Each of the package shall be marked on three sides with indelible paint of proper quality: indicating contract number and date, brief description of goods including quantity, Packing list reference number, country of origin of goods and any other relevant details.
10. **Spare Parts:** Seller shall provide materials, information etc. pertaining to spare parts manufactured and supplied by the OEM. It shall be ensured that the required spares are available for purchase at least for 10 years from date of supplies. In case due to any reasons the production of the spare parts is discontinued sufficient advance notice should be given to the buyer/consignee before such discontinuation to provide adequate time to purchase the required spare parts etc. Further, OEM and their service centres/dealers shall carry sufficient inventories to assure ex-stock supply of consumables and spares for the equipments so that the same are available. OEM or reseller shall always accord most favoured client status to the buyer/consignee and shall give the most competitive price for spares and consumables of its machines/equipments supplied.
11. **Installation, Training, Manuals:** Seller shall be responsible to carry out Installation & commissioning, Supervision and Demonstration of the goods. They shall provide required jigs and tools for assembly, minor civil works for the completion of the installation and Training of Consignee's representatives for operating and maintaining the equipment and supplying required number of operation & maintenance manual for the goods. In case the category parameters are specifying any requirements regarding the installations, training and manuals the same shall also be applicable.
12. **Electrical safety checking:** Sellers are required to make sure that they furnish the list of equipments for carrying out routine and preventive maintenance to buyer/consignee .They should make sure to periodically check the electrical safety aspects as per BIS Safety Standards or equivalent .In case they do not have required equipment for such testing should ensure that the equipments checked for electrical safety compliance through labs with facilities for such checking during every preventive maintenance call.
13. **Software:** All software updates should be provided free of cost during warranty period.

क्रेता द्वारा जोड़ी गई बिड की विशेष शर्तें/Buyer Added Bid Specific Terms and Conditions

1. Experience Certificate for the supply of the same to any Govt/ PSU/ any renowned private organisation along with Supply/ Purchase Order.
2. If the agency is registered under MSME or NSIC, then EMD exemption certificate needs to be enclosed.
3. Make in india specific authorisation certificate needs to be enclosed.

अस्वीकरण/Disclaimer

The additional terms and conditions have been incorporated by the Buyer after approval of the Competent Authority in Buyer Organization, whereby Buyer organization is solely responsible for the impact of these clauses on the bidding process, its outcome, and consequences thereof including any eccentricity / restriction arising in the bidding process due to these ATCs and due to modification of technical specifications and / or terms and conditions governing the bid. If any clause(s) is / are incorporated by the Buyer regarding following, the bid and resultant contracts shall be treated as null and void and such bids may be cancelled by GeM at any stage of bidding process without any notice:-

1. Definition of Class I and Class II suppliers in the bid not in line with the extant Order / Office Memorandum issued by DPIIT in this regard.
2. Seeking EMD submission from bidder(s), including via Additional Terms & Conditions, in contravention to exemption provided to such sellers under GeM GTC.

3. Publishing Custom / BOQ bids for items for which regular GeM categories are available without any Category item bunched with it.
4. Creating BoQ bid for single item.
5. Mentioning specific Brand or Make or Model or Manufacturer or Dealer name.
6. Mandating submission of documents in physical form as a pre-requisite to qualify bidders.
7. Floating / creation of work contracts as Custom Bids in Services.
8. Seeking sample with bid or approval of samples during bid evaluation process. (However, in bids for attached categories, trials are allowed as per approved procurement policy of the buyer nodal Ministries)
9. Mandating foreign / international certifications even in case of existence of Indian Standards without specifying equivalent Indian Certification / standards.
10. Seeking experience from specific organization / department / institute only or from foreign / export experience.
11. Creating bid for items from irrelevant categories.
12. Incorporating any clause against the MSME policy and Preference to Make in India Policy.
13. Reference of conditions published on any external site or reference to external documents/clauses.
14. Asking for any Tender fee / Bid Participation fee / Auction fee in case of Bids / Forward Auction, as the case may be.
15. Buyer added ATC Clauses which are in contravention of clauses defined by buyer in system generated bid template as indicated above in the Bid Details section, EMD Detail, ePBG Detail and MII and MSE Purchase Preference sections of the bid, unless otherwise allowed by GeM GTC.
16. In a category based bid, adding additional items, through buyer added additional scope of work/ additional terms and conditions/or any other document. If buyer needs more items along with the main item, the same must be added through bunching category based items or by bunching custom catalogs or bunching a BoQ with the main category based item, the same must not be done through ATC or Scope of Work.

Further, if any seller has any objection/grievance against these additional clauses or otherwise on any aspect of this bid, they can raise their representation against the same by using the Representation window provided in the bid details field in Seller dashboard after logging in as a seller within 4 days of bid publication on GeM. Buyer is duty bound to reply to all such representations and would not be allowed to open bids if he fails to reply to such representations.

Additional Clause For Comprehensive Maintenance Charges

- 1.CMC shall include preventive maintenance including calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, after satisfactory completion of Warranty. During the CMC period commencing from date of the successful completion of warranty period, Service personnel shall visit each consignee site as recommended in the manufacturer's technical/ service /operational manual, at least once in six months or as per user requirement. Cost of consumables shall not be included in CMC.Further there will be 98% uptime warranty during CMC period on 24 (hrs) X 7 (days) X 365 (days) basis, with penalty, to extend CMC period by double the downtime period.
- 2.CMC charges to be indicated as percentage of cost of equipment quoted for each year after the warranty period.
- 3.GST shall be included in the CMC Charges quoted.
- 4.Cost of CMC will be added for Ranking/Evaluation purpose with depreciation formula.A 10% discounting rate per year shall be applied on CMC Charges for price evaluation on Net Present Value.
- 5.The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user.
- 6.While creating a bid or RA, buyers shall indicate whether CMC is required against Yes/No" options. If CMC Charges are included, an option for number of years for CMC required after the warranty period shall be available.Under this option up to 10 years can be chosen for CMC charges beyond warranty period.
- 7.In case the bid has a provision for CMC, the warranty of the product will also be deemed to have been converted into Comprehensive warranty including preventive maintenance and calibration as per technical/ service /operational manual of the manufacturer, service charges and spares, during the Warranty Period also. Sellers are therefore advised to include the cost of Comprehensive Warranty including spares (excluding consumables) also in product Cost.
- 8.The CMC functionality shall be available in bid only and no direct RA shall be applicable.In case of bid to R/A decrement rules shall be applicable on total price inclusive of CMC charges. Bunching of products shall not be available while creating bids with CMC charges.

8.1.Buyer shall indicate number of years of warranty by selecting different options available in the field depending on warranty parameter applicable in category parameters for the equipment. No. of years of warranty indicated here shall supersede the warranty period indicated elsewhere in bid or product specifications. The Seller while participating in Bid/RA will get fields to indicate CMC charges as percentage depending on number of years of CMC selected by Buyer. The following shall be applicable, if 5 year CMC selected:

CMC charges for 1st year after warranty period- Percentage to be indicated- A1

CMC charges for 2nd year after warranty period- Percentage to be indicated- A2

CMC charges for 3rd year after warranty period - Percentage to be indicated- A3

CMC charges for 4th year after warranty period - Percentage to be indicated- A4

CMC charges for 5th year after warranty period - Percentage to be indicated- A5

Similarly, A6 to A10 are to be indicated for 6th to 10th year of CMC if applicable.

8.2.The calculation of CMC Charges shall take into account the number of years of warranty and duration of CMC as specified while creating bid.

8.3.In the price evaluation, the system shall provide function to calculate the cost of each equipment by formula indicated below including CMC and then show the inter-se-ranking of the bidders. The following are the variables

(i) Number of years for which CMC required.

(ii) Number of years of product warranty

The formula for calculating total cost including CMC charges shall be as under:

Total Cost for evaluation=

$C+C^* \{(A1/100)/(1.10^n)+(A2/100)/(1.10^{n+1})+(A3/100)/(1.10^{n+2})+(A4/100)/(1.10^{n+3})+(A5/100)/(1.10^{n+4})\}$ and so on

C - Cost for equipment quoted and n shall be number of years of product warranty specified.

If 2 year warranty specified, n shall be 2 and if 5 year warranty specified, n shall be 5. A1, A2, A3, A4 & A5 shall depend on how many years CMC selected. For 3 year CMC, only A1, A2 and A3 factors are to be taken into account and A4 and A5 will not be applicable.

8.4.CMC charges offered for each subsequent year should be same or higher than preceding year.

8.5.The CMC charges shall be offered within range of 3 to 25% of cost of equipment.

9.Since CMC charges are to be paid only later for each year during CMC period, applicable performance guarantee amount after placement of contract shall be based on the cost of equipment excluding the cost of CMC Charges.

10.Performance bank guarantee applicable for CMC is to be submitted at start of the CMC and shall be applicable between 2.5% to 10% as specified in bid on total CMC Charges. The PBG submitted after award of contract shall be released only after new PBG for the CMC period is submitted and accepted by buyer/consignee after due verification. Bank guarantee for CMC is to remain valid till completion of CMC period plus one year. The bank guarantee for CMC shall be submitted to buyer directly. In case, seller fails to submit the PBG or does not provide services for the CMC contract after expiry of warranty period then PBG of equipment shall be forfeited.

11.In case of splitting of order quantity, equipment cost and CMC charges offered by L1 bidder shall be matched by higher quoting eligible bidders on one-to-one basis. The equipment cost and CMC charges (year to year) shall be matched individually.

12.The CMC Contract shall be an offline contract to be handled by buyer. The payment of CMC will be made on quarterly basis after satisfactory completion of said period, duly certified by end user and scope of CMC will be as per para 1 above.

13.CMC Charges are inclusive of all the charges for Transportation, Lodging, Boarding, all insurances including third party insurance and all other incidental charges. The same shall include GST. The prices also include cost of spares and damaged parts. Purchaser does not have any liability, whatsoever, over and above the cost of CMC. It also includes for arranging hand tools & tackles, special tools etc. required to carry out the work.

All GeM Sellers / Service Providers are mandated to ensure compliance with all the applicable laws / acts / rules including but not limited to all Labour Laws such as The Minimum Wages Act, 1948, The Payment of Wages Act, 1936, The Payment of Bonus Act, 1965, The Equal Remuneration Act, 1976, The Payment of Gratuity Act, 1972 etc. Any non-compliance will be treated as breach of contract and Buyer may take suitable actions as per GeM Contract.

यह विड सामान्य शर्तों के अंतर्गत भी शासित है / This Bid is also governed by the General Terms and Conditions

जेम की सामान्य शर्तों के खंड 26 के संदर्भ में भारत के साथ भूमि सीमा साझा करने वाले देश के बिडर से खरीद पर प्रतिबंध के संबंध में भारत के साथ भूमि सीमा साझा करने वाले देश का कोई भी बिडर इस नियिदा में बिड देने के लिए तभी पात्र होगा जब वह बिड देने वाला सक्षम प्राधिकारी के पास पंजीकृत हो। बिड में भाग लेते समय बिडर को इसका अनुपालन करना होगा और कोई भी गलत घोषणा किए जाने व इसका अनुपालन न करने पर अनुबंध को तत्काल समाप्त करने और कानून के अनुसार आगे की कानूनी कार्रवाई का आधार होगा। /In terms

of GeM GTC clause 26 regarding Restrictions on procurement from a bidder of a country which shares a land border with India, any bidder from a country which shares a land border with India will be eligible to bid in this tender only if the bidder is registered with the Competent Authority. While participating in bid, Bidder has to undertake compliance of this and any false declaration and non-compliance of this would be a ground for immediate termination of the contract and further legal action in accordance with the laws.

---धन्यवाद/Thank You---